



Republic of the Philippines
 Department of Agriculture
Philippine Center for Postharvest Development and Mechanization
 Science City of Muñoz, Nueva Ecija
www.philmech.gov.ph
 ISO 9001:2015, ISO 14001:2015, ISO 45001:2018 CERTIFIED



BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL BID BULLETIN NO. 1

**SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT
 (PHilMech-Goods-23-05-10)**

This **Supplemental/Bid Bulletin No. 1** is being issued to clarify, amend and supplement provisions in the Bidding Documents.

A. ISSUES AND CONCERNS DURING THE PRE-BID CONFERENCE

1. <i>Issue / Proposal: Request for the wavelength range of the Microplate Reader to be monochromator.</i>		
Action Taken/Reply: Technical Specification is retained.		
2. <i>Issue / Proposal: Request for the printer to be compatible with a computer or laptop.</i>		
Action Taken/Reply: Request/Proposal is granted. Technical Specification is amended.		
	FROM	TO
Technical Specifications (LOT 2)	Microplate Reader B.2 PRINTER 3-in-1 Printer (print, scan, copy) compatible with the Microplate Reader with 4 color-ink; continuous ink-tank system with two (2) extra sets of ink (set composed of Black, Magenta, Cyan and Yellow)	Microplate Reader B.2 PRINTER 3-in-1 Printer (print, scan, copy) compatible with the Computer or Laptop with 4 color-ink; continuous ink-tank system with two (2) extra sets of ink (set composed of Black, Magenta, Cyan and Yellow)
3. <i>Issue / Proposal: Inquiry on the speed range of Rotary Shaker – at least 10 - 250 rpm or better.</i>		
Action Taken/Reply: The speed range can go lower than 10rpm and higher than 250rpm, but not higher than 10rpm nor lower than 250rpm.		
4. <i>Issue / Proposal: Request for the timer of the Rotary Shaker to be at least 50 hr 59 min or lower.</i>		
Action Taken/Reply: Technical Specification is retained.		

B. OTHER AMENDMENTS

	FROM	TO
Technical Specifications (Lot 1)	Sample Concentrator H. OTHER REQUIREMENTS H.5 The supplier must have at least five (3) local installations of Sample Concentrator that can be viewed and demo to the end user, if necessary. One (1) out of three (3) local installations must be the same model as the one being offered.	Sample Concentrator H. OTHER REQUIREMENTS H.5 The supplier must have at least three (3) local installations of Sample Concentrator that can be viewed and demo to the end user, if necessary. One (1) out of three (3) local installations must be the same model as the one being offered.


Main Office: CLSU, Science City of Muñoz, Nueva Ecija, Philippines
 Mobile Nos.: (+63) 932-869-6837 / (+63) 917-813-0852
 Fax No.: (044) 456-0110
Liaison Office: 3rd Floor, ATI Bldg., Elliptical Road, Diliman, Quezon City
 Tel No.: (02) 927-4019, 927-4029




Technical Specifications (Lot 4)	Refrigerated (or Cooled) Incubator WARRANTY AND AFTER SALES SERVICES Note: Commitment Letter must be provided upon submission of bid	
	G. TRAINING Note: Commitment Letter must be provided upon submission of bid	
	H. OTHER REQUIREMENTS Note: Commitment Letter must be provided upon submission of bid	
Technical Specifications (Lot 5)	Rotary Shaker H. OTHER REQUIREMENTS Note: Commitment Letter must be provided upon submission of bid	


Issued this 17th day of July 2023 in the Science City of Muñoz, Nueva Ecija, Philippines.


Reviewed and approved by:


RONALDO SEBASTIAN R. REYES
Chairperson, Bids and Awards Committee

ORIGINAL SIGNED
OFERO A. CAPARIÑO, Ph.D
Vice Chairperson


NELSON C. SANTIAGO, Ph.D
Regular Member


ELEONOR F. SANTIAGO
Technical Member


CLARISSA C. JUAN
End-User/Provisional Member

CONFORMED: _____
Bidders Name: _____
Date Received: _____