

Republic of the Philippines Department of Agriculture Philippine Center for Postharvest Development and Mechanization Science City of Muñoz, Nueva Ecija, Philippines www.philmech.gov.ph

Negotiation Documents

(Negotiation - Two-Failed Biddings Pursuant to Sec. 53.1 of 2016 Revised IRR of RA 9184)

SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT

(PHilMech-Goods-23-11-29-NP)

Schedule of Activities:

Pre-Negotiation Conference: **November 30, 2023 ; 1:30PM** Auditorium, PHilMech Main Office Science City of Muñoz, Nueva Ecija

Submission & Opening of Best and Final Offer: **December 13, 2023 ; 3:00PM** Auditorium, PHilMech Main Office, Science City of Muñoz, Nueva Ecija

> Sixth Edition July 2020

Table of Contents

Gloss	ary of Acronyms, Terms, and Abbreviations	<u>3</u>
Sectio	on I. Invitation for Negotiation	6
	on II. Instructions to Bidders	
1.	Scope of Bid	10
2.	Funding Information	10
3.	Requirements for Negotiated Procurement	10
4.	Corrupt, Fraudulent, Collusive, and Coercive Practices	10
5.	Eligible Bidders	10
6.	Origin of Goods	11
7.	Subcontracts	11
8.	Pre-Negotiation Conference	11
9.	Checklist of Requirements	
	Error! Bookmark not defined.	
10.	Bid Prices	
11.	Bid and Payment Currencies	12
12.	Bid Security	
13.	Sealing and Marking of Bids	
14.	Deadline for Submission of Bids	
15.	Opening and Preliminary Examination of Bids	
16.	Domestic Preference	13
17.	Detailed Evaluation and Comparison of Bids	
18.	Post-Qualification	
19.	Signing of the Contract	14
Sectio	on III. General Conditions of Contract	15
1.	Scope of Contract	16
2.	Advance Payment and Terms of Payment	16
3.	Performance Security	16
4.	Inspection and Tests	16
5.	Warranty	17
6.	Liability of the Supplier	17
Sectio	on IV. Special Conditions of Contract	18
Sectio	on V. Schedule of Requirements	20
	on VI. Technical Specifications	
	on VII. Checklist of Technical and Financial Documents	

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC – Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP - Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP - Refers to the quoted price of the Goods, which means "delivered duty paid."

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – "Free Carrier" shipping point.

FOB - "Free on Board" shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC - Government-owned and/or -controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB - Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national

buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Section I. Invitation for Negotiation



Republic of the Philippines DEPARTMENT OF AGRICULTURE PHILIPPINE CENTER FOR POSTHARVEST DEVELOPMENT AND MECHANIZATION ISO 9001:2015, ISO 14001:2015, ISO 45001:2018 CERTIFIED www.philmech.gov.ph



INVITATION FOR NEGOTIATION

SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT

(PHilMech-Goods-23-11-29-NP)

1. The **Philippine Center for Postharvest Development and Mechanization (PHilMech)** through its **FY 2023 Regular Fund** intends to apply the below-specified amount, being the Approved Budget for the Contract (ABC), as indicated herein to payments under the contract for the **SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT.** Bids received in excess of the ABC shall be automatically rejected during bid opening.

LOT Number	PARTICULARS	ABC (PhP)	COST OF BIDDING DOCUMENTS (PhP)
1	Supply, Delivery, Installation, Commissioning and Testing of 1 Sample Concentrator complete with all necessary parts and accessories	200,000.00	200.00
2	Supply, Delivery, Installation, Commissioning and Testing of 1 Incubator complete with all necessary parts and accessories	300,000.00	300.00
3	Supply, Delivery, Installation, Commissioning and Testing of 1 Refrigerated (or Cooled) Incubator complete with all necessary parts and accessories	600,000.00	600.00
4	Supply, Delivery, Installation, Commissioning and Testing of 1 Rotary Shaker complete with all necessary parts and accessories	150,000.00	200.00
ALL LOT	rs	1,250,000.00	1,300.00

- 2. The **PHilMech** now invites bids from eligible bidders for the **SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT**. Delivery of the GOODS is required within **Sixty (60) calendar days for Lot I, Ninety (90) calendar days for Lots II & IV and One Hundred Twenty (120) calendar days for Lot III** from receipt of the Notice to Proceed. Bidders should have completed within ten (10) years from the date of submission and receipt of bids a contract similar to the project. For this purpose, a similar contract shall refer to any contract for the **Supply and Delivery of Laboratory Equipment**.
- 3. Bidding will be conducted through **Negotiated Procurement (Two-Failed Biddings)** pursuant to Section 53.1 of the Revised Implementing Rules and Regulations (IRR) of the Republic Act 9184, otherwise known as the "Government Procurement Reform Act", including its Annex H.

The BAC will engage in negotiations with a sufficient number of suppliers to ensure effective competition. The selection of the successful offer shall be based on the best and final offer that will be submitted on the specified date shown below, and which would meet the minimum technical specifications required.

Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183.

4. Interested Bidders may obtain further information from the **PHilMech** and inspect the Negotiation Documents at the address given below from 8:00 am to 5:00 pm.

5. A complete set of Negotiation Documents may be acquired by interested Bidders during office hours from **November 28 to December 13, 2023** from the address given below upon payment of the applicable fee for the Negotiation Documents as specified above.

Bidders may also opt to deposit in cash the applicable fee to the PHilMech's Official Bank Account: Landbank of the Philippines (LBP) Account Name: <u>PHILMECH TRUST FUND</u>; Account No. 2962-1023-63 and by presenting or sending the proof of payment e.g. bank deposit slips to the address written below or email to <u>bac@philmech.gov.ph</u> subject to the condition that the bank service charge will be shouldered by the bidder.

The Negotiation Documents may also be downloaded free of charge from the website of the Philippine Government Electronic Procurement System (PhilGEPS) and the website of the Procuring Entity provided that Bidders shall pay the applicable fee for the Bidding Documents before the submission of their Bids.

6. The PHilMech will hold a Negotiation Conference on November 30, 2023 ; 1:30PM at the Auditorium, PHilMech Main Office, CLSU Compound, Science City of Muñoz, Nueva Ecija, which shall be open to prospective bidders.

Interested bidders may opt to attend via video conferencing and pre-register on the email address provided below to signify their intention to participate. A Personal Meeting ID shall be provided to the authorized representative upon registration.

- 7. Best and final quotations must be duly received by the BAC Secretariat through manual submission at the address indicated below on or before **3:00PM of December 13, 2023.** Late Bids shall not be accepted.
- 8. Bid opening shall be on **December 13, 2023 ; 3:00PM** at the address given below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
- 10. The **PHilMech** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Section 35.6 and Section 41 of RA9184 and its IRR and GPPB Resolution No. 09-2020 without thereby incurring any liability to the affected bidder or bidders.
- 11. For further information, please contact: The BAC Secretariat, PHilMech Administrative Division-Procurement Management Section CLSU Compound, Science City of Muñoz, Nueva Ecija Contact No. 0917 301 8242 E-mail: <u>bac@philmech.gov.ph</u>
- 12. Interested bidders may also visit the following websites: For downloading of Bidding Documents: www.philmech.gov.ph https://old.philmech.gov.ph/?page=bidding

NELSON C. SANTIAGO, Ph.D. BAC Chairperson

NOTICE/ANNOUNCEMENT

PHILMECH DOES NOT CONDONE ANY FORM OF SOLICITATION FROM ANY PROSPECTIVE BIDDER BY ANY OF ITS STAFF/EMPLOYEE OR BY ANY OTHER PARTY. ANY ACTIVITY OF THIS NATURE MAY BE IMMEDIATELY REPORTED TO THE OFFICE OF THE DIRECTOR OR TO ANY APPROPRIATE LAW ENFORCEMENT AGENCY FOR ENTRAPMENT AND/OR INVESTIGATION.

Section II. Instructions to Bidders

1. Scope of Bid

The Procuring Entity, *Philippine Center for Postharvest Development and Mechanization (PHilMech)* wishes to receive Bids for the *SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT*, with identification number (*PHilMech-Goods-23-11-29-NP*).

The Procurement Project SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT is composed of *four (4) lots*, the details of which are described in Section VII (Technical Specifications)

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for FY 2023 in the amount of Php 1,250,000.00.
- 2.2. The source of funding is NGA, the General Appropriations Act or Special Appropriations.

3. Requirements for Negotiated Procurement

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or IN by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.

- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

The Procuring Entity has prescribed that Subcontracting is not allowed.

8. Pre-Negotiation Conference

The Procuring Entity will hold a negotiation conference for this Project on the specified date and time and either at the PHilMech Auditorium and/or video conferencing/webcasting as indicated in paragraph 6 of the IN.

9. Checklist of Requirements

□ (a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;

- (b) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
 <u>or</u> Original copy of Notarized Bid Securing Declaration; <u>and</u>
- (c) Original of duly signed and accomplished Quotation Form

10. Bid Prices

- 11.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.
 - b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in Section VII (Technical Specifications).

11. Bid and Payment Currencies

- 12.1 For Goods that the Bidder will supply from outside the Philippines, the bid prices shal be quoted in the local currency.
- 12.2 Payment of the contract price shall be made in Philippine Pesos.

12. Bid Security

- 13.1 The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security allowed under Section 27.2 of the 2016 Revised Implementing Rules and Regulations.
- 13.2 The Bid and bid security shall be valid until *April 11, 2024 or one hundred twenty (120) calendar days after the opening of bids in the event of a change in the date of the opening of bids.* Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

13. Sealing and Marking of Bids

Each Bidder shall submit three (3) copies of the first and second components of its Bid.

14. Deadline for Submission of Bids

The Bidders shall submit its best and final offer on the specified date and time as indicated in paragraph 7 of the IN.

15. Opening and Preliminary Examination of Bids

16.1 The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IN**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

16.2 The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

16. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

17. Detailed Evaluation and Comparison of Bids

- 19.2. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.3. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.4. The descriptions of the lots or items shall be indicated in **Section VII** (**Technical Specifications**), although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.5. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.6. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

18. Post-Qualification

20.2 Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

19. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. General Conditions of Contract

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC).**

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC**, **Section VII** (**Technical Specifications**) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

- 5.1 In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 5.2 The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section IV. Special Conditions of Contract

Special Conditions of Contract

GCC	
Clause	
1	Delivery and Documents – For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:
	Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements) and Terms of Reference (TOR).
	For purposes of this Clause the Procuring Entity's Representative of the project is Ms. Clarissa Juan, LSD Senior Science Research Specialist for Lots I & II and Mr. Angelo A. Dela Fuente, FPD Science Research Specialist II for Lots III & IV.
2.2	The terms of payment shall be as follows:
	A. One Hundred percent (100%) of the Contract Price: upon full delivery, inspection and acceptance
4	All goods delivered under this contract shall be subject to the following test:
	 Operational / Performance Test Technical inspection and acceptance
5.1	The Procuring Entity's address for notice is philmech_bac@yahoo.com.ph The Bidder's address for notice is the e-mail address indicated in the bidder's Omnibus Sworn Statement (OSS) Service of pertinent notices related to this bidding through Electronic Mail (e- mail) to the email address provided by the bidder in the OSS shall be
	tantamount to proof of actual service and shall be deemed received by the bidder on the date of service or email.

Section V. Schedule of Requirements

RE: SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT

(PHilMech-Goods-23-11-29-NP)

Lot no/s. being bid:	Name and Signature of Authorized Representative:

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number (LOT No.)	PARTICULARS	Delivered, Weeks/ Months
1	Supply, Delivery, Installation, Commissioning and Testing of 1 Sample Concentrator complete with all necessary parts and accessories	60CD
2	Supply, Delivery, Installation, Commissioning and Testing of 1 Incubator complete with all necessary parts and accessories	90CD
3	Supply, Delivery, Installation, Commissioning and Testing of 1 Refrigerated (or Cooled) Incubator complete with all necessary parts and accessories	120CD
4	Supply, Delivery, Installation, Commissioning and Testing of 1 Rotary Shaker complete with all necessary parts and accessories	90CD

Signature over Printed Name of Bidder

Position/Designation

(Name, Address & Tel. No. of Company)

Date

Technical Specifications

RE: SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT

(PHilMech-Goods-23-11-29-NP)

Name and Signature of Authorized Representative:

		Statement of Compliance
	TFICATION LOT 1)	Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.
Supply, Delivery, Installation, Com		
Concentrator complete with all nece	essary parts and accessories	
A.1 Technical Specification		
Parameters	Specifications	
Temperature Setting/Control	ambient +5°C to 100°C or wider range	
Temperature Display Accuracy	≤ 0.5 °C or better	
Display	Digital (LED or LCD)	
Time Range/Control	at least 1min to 10h 59min or better	
Test Tube Diameter	must include 13-19 mm outside	
N:	diameter or wider range	
Nitrogen flowrate	at least 0 to 10 L/min or better	
Sample Capacity	at least 24 12 or higher	
Gas Needle Control	Combination or Individual control	

``Read and accepted as part of the contract"

B. ACCESORIES, SPARES AND CONSUMABLES	
B.1 Any other accessories require to be quoted to make the system fully operational must be included in the offer.	
C. POWER SUPPLY	
C.1 Power input 220V-230 V or Auto volt AC, 50 - 60 Hertz or wider range	
C.2 Automatic Voltage Regulator (AVR) compatible with the unit	
D. WARRANTY, SERVICE AND AFTER SALES	
D.1 A minimum of two (2) years warranty on parts and services of Sample Concentrator. Commencing after successful installation by Supplier and its acceptance by PHilMech	
D.2 During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.	
D.3 Provision of annual preventive maintenance (PM) for a period of two (2) years for the Sample Concentrator, after acceptance by PHilMech. The PM must include at minimum the cleaning, testing and checking of various functions.	
D.4 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable	
D.5 Instrument should be attended for repair by trained service personnel within 10 working days from the time fault is reported by the user.	
The equipment shall be made fully operational after 10 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (11 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.6 In case there is a transfer of sole-distributor rights, the former distributor	
shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE	
 E. TESTING AND ACCEPTANCE E.1 Acceptance testing will occur after completion of the training course and after the supplier states the system is installed and ready to perform in compliance with specifications F. DOCUMENTATION 	
F.1 Training certificate of Service Personnel from the manufacturer related to the installation, equipment maintenance and troubleshooting must be provided.	
F.2 Manufacturer's certificate of compliance and inspection must be provided.	
F.3 One (1) original copy of operating and service manuals in English.	
F.4 IQ, OQ, and PQ protocols whichever is applicable. The protocols must provide the necessary information and test procedures required to prove that the instrument meets the specified requirements.	
 F.5 Submission of IQ, OQ, and PQ (whichever is applicable) reports within five (5) working days after completion of qualification. Certificate of qualification must also be issued, if applicable. G. TRAINING 	
G.1 The training should cover lectures on the theory and principles of the instrument. Troubleshooting, maintenance and service of equipment's performance should also be discussed. Lecture and hands-on training should be conducted at the PHilMech Office for at least five (5) participants. The training shall be provided at no cost to the PHilMech	
H. OTHER REQUIREMENTS	
H.1 The brand has been in the market for at least five (5) years.	

H.2 Supplier must submit a proof of authorization from the Manufacturer of the offered equipment. Authorization must include sales, distribution of equipment and its parts, and after sales services, at a minimum.	
H.3 Provide certification of availability and continuity of spare parts for at	
least 5 years after delivery	
I. DELIVERY PERIOD	
I.1 Not more than sixty (60) calendar days upon receipt of Notice to Proceed.	
To be delivered at Laboratory Services Division, PHilMech, CLSU Compound,	
Science City of Muñoz, Nueva Ecija.	

I hereby certify that the Statement of Compliance to the Foregoing Technical Specifications are true and correct, otherwise, if found false either during the bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Signature over Printed Name of Bidder

Position / Designation

(Name, Address & Tel. No. of Company)

		Statement of Compliance
	SPECIFICATION (LOT 2)	Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.
Supply, Delivery, Installation complete with all necessary	n, Commissioning and Testing of Incubator parts and accessories	
	L · · · · · · · · · · · · · · · · · · ·	
A.1 Technical Specification		
A.1 Technical Specification Parameters	Specifications	
A.1 Technical Specification	Specifications Width – at least 600 to 778mm	
A.1 Technical Specification Parameters		
A.1 Technical Specification Parameters	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm	
A.1 Technical Specification Parameters	Width – at least 600 to 778mmHeight – at least 1000 to 1900 mm	
A.1 Technical Specification Parameters Exterior Dimension	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity	Vidth - at least 600 to 778mmHeight - at least 1000 to 1900 mmDepth - 600 to 800 mmAmbient +5°C to 60°C or wider ≥ 0.5 °C or betterat least 350 - 400 liters or higher (compatible with the exterior dimension)	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity	Vidth - at least 600 to 778mmHeight - at least 1000 to 1900 mmDepth - 600 to 800 mmAmbient +5°C to 60°C or wider ≥ 0.5 °C or betterat least 350 - 400 liters or higher (compatible with the exterior dimension)	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s	Width – at least 600 to 778mm Height – at least 1000 to 1900 mm Depth – 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s No. of Inner glass door/s	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1 at least 1	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s No. of Inner glass door/s No. of racks/shelf	Width - at least 600 to 778mm Height - at least 1000 to 1900 mm Depth - 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1 at least 1 at least 2, compatible with the incubator unit Include audible or visual alarm system for errors	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s No. of Inner glass door/s No. of racks/shelf Monitor and Alarm B. ACCESORIES, SPARES AN	Width – at least 600 to 778mm Height – at least 1000 to 1900 mm Depth – 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1 at least 1 at least 2, compatible with the incubator unit Include audible or visual alarm system for errors ND CONSUMABLES re to be quoted to make the system fully	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s No. of Inner glass door/s No. of racks/shelf Monitor and Alarm B. ACCESORIES, SPARES AN Any other accessories requin	Width – at least 600 to 778mm Height – at least 1000 to 1900 mm Depth – 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1 at least 1 at least 2, compatible with the incubator unit Include audible or visual alarm system for errors ND CONSUMABLES re to be quoted to make the system fully	
A.1 Technical Specification Parameters Exterior Dimension Temperature Range Resolution Volume Capacity Display No. of External door/s No. of Inner glass door/s No. of racks/shelf Monitor and Alarm B. ACCESORIES, SPARES AN Any other accessories requin operational must be included C. POWER SUPPLY C.1 Power input 220V-230 V	Width – at least 600 to 778mm Height – at least 1000 to 1900 mm Depth – 600 to 800 mm Ambient +5°C to 60°C or wider ≥ 0.5°C or better at least 350 - 400 liters or higher (compatible with the exterior dimension) Display: Digital display (LED or LCD) or Vacuum Fluorescent Display at least 1 at least 1 at least 2, compatible with the incubator unit Include audible or visual alarm system for errors ND CONSUMABLES re to be quoted to make the system fully	

``Read and accepted as part of the contract"

WARRANTY AND AFTER SALES SERVICES	
D.1 A minimum of two (2) years warranty on parts and services of Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech	
During the warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.	
D.2 Provision of annual preventive maintenance (PM) for a period of two (2) years for the Incubator system, after acceptance by PHilMech. The PM must include at minimum the cleaning, testing and checking of various functions	
D.3 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ), Performance Qualification (PQ) and Verification/Validation Qualification, whichever is applicable	
D.4 Instrument should be attended for repair by trained service personnel within 10 working days from the time fault is reported by the user.	
The equipment shall be made fully operational after 10 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (11 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation.	
D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE	
E.1 Acceptance testing will occur after completion of the training course and after the supplier states the system is installed and ready to perform in compliance with specifications	
E.2 Includes calibration by ISO 17025 Calibration Service Provider prior to acceptance with 3 set points set by the end-user (with Calibration Certificate) F. DOCUMENTATION	
F.1 Training certificate of Service Personnel from the manufacturer related to the installation, equipment maintenance and troubleshooting must be provided.	
F.2 Manufacturer's certificate of compliance and inspection must be provided.	
F.3 One (1) original copy of operating and service manuals in English.	
F.4 IQ, OQ, PQ and/or Verification/Validation Qualification protocols whichever is applicable. The protocols must provide the necessary information and test procedures required to prove that the instrument meets the specified requirements.	
F.5 Submission of IQ, OQ, PQ and/or Verification/Validation qualification (whichever is applicable) reports within five (5) working days after completion of qualification. Test data (raw data) and calibration certificates of the calibrating instruments must be attached. Certificate of qualification must also be issued, if applicable.	
G. TRAINING	
G.1 The training should cover lectures on the theory and principles of the instrument and its applications. Troubleshooting, maintenance and service of equipment's performance should also be discussed. Lecture and hands-on training should be conducted at the PHilMech Office for at least five (5)	
participants. The training shall be provided at no cost to the PHilMech H. OTHER REQUIREMENTS	
H.1 The brand has been in the market for at least five (5) years.	
H.2 Supplier must submit a proof of authorization from the Manufacturer of the offered equipment. Authorization must include sales, distribution of	

equipment and its parts, and after sales services, at a minimum.	
H.3 Provide certification of availability and continuity of spare parts for at least 5 years after delivery	
H.4 The supplier must have at least two (2) local installations of Incubator.	
Note: Upon submission of bid, the bidder must submit a list of companies with installed units. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted. I. DELIVERY PERIOD	
I.1 Not more than ninety (90) calendar days upon receipt of Notice to Proceed.	
To be delivered at Laboratory Services Division, PHilMech, CLSU Compound,	
Science City of Muñoz, Nueva Ecija.	

I hereby certify that the Statement of Compliance to the Foregoing Technical Specifications are true and correct, otherwise, if found false either during the bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Signature over Printed Name of Bidder

Position / Designation

(Name, Address & Tel. No. of Company)

		Statement of Compliance
SPECIFICATION (LOT 3) Supply, Delivery, Installation, Commissioning and Testing of Refrigerated (or Cooled) Incubator complete with all necessary parts and		Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross- referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.
A.1 Technical Specification	on	
Parameters	Specifications	
Parameters Incubator Type	Specifications with Biochemical Oxygen Demand (BOD) system	
Parameters	Specificationswith Biochemical Oxygen Demand (BOD)systemWidth - 650 to 900mm	
Parameters Incubator Type	Specifications with Biochemical Oxygen Demand (BOD) system	
Parameters Incubator Type Exterior Dimension	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mm	
Parameters Incubator Type	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mm	
Parameters Incubator Type Exterior Dimension	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mm	
Parameters Incubator Type Exterior Dimension Interior Dimension	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm	
Parameters Incubator Type Exterior Dimension Interior Dimension Temperature Range	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm $\leq +20^{\circ}$ C to $\geq 50^{\circ}$ C or wider	
Parameters Incubator Type Exterior Dimension Interior Dimension Temperature Range Resolution	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm $\leq +20^{\circ}$ C to $\geq 50^{\circ}$ C or wider $\leq 0.1^{\circ}$ C or better	
Parameters Incubator Type Exterior Dimension Interior Dimension Temperature Range	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm $\leq +20^{\circ}$ C to $\geq 50^{\circ}$ C or wider	
Parameters Incubator Type Exterior Dimension Interior Dimension Temperature Range Resolution Volume Capacity Display	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm $\leq +20^{\circ}$ C to $\geq 50^{\circ}$ C or wider $\leq 0.1^{\circ}$ C or betterat least 450 to 600 liters or higherDigital display (LED or LCD)	
Parameters Incubator Type Exterior Dimension Interior Dimension Temperature Range Resolution Volume Capacity	Specificationswith Biochemical Oxygen Demand (BOD) systemWidth - 650 to 900mmHeight - 1800 to 2000 mmDepth - 600 to 800 mmWidth - 500 to 700mmHeight - 1200 to 1600 mmDepth - 500 to 630 mm $\leq +20^{\circ}$ C to $\geq 50^{\circ}$ C or wider $\leq 0.1^{\circ}$ C or betterat least 450 to 600 liters or higher	

``Read and accepted as part of the contract"

No. of racks/shelf 4 - 6 compatible with the incubator unit Monitor and Alarm Include audible or visual alarm system for errors and fault analysis Environmental operating conditions Philippine environmental conditions: at least B. ACCESORIES, SPARES AND CONSUMBLES Any other accessories require to be quoted to make the system fully operational must be included in the offer. C. POWER SUPPLY C.1 Power input should be 220V - 230V, 60Hz C.2 Automatic Voltage Regulator (AVR) compatible with the unit WARANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (1Q), Operational Qualification (0Q) and Performance Qualification (1Q), Operational Qualification package. Package must include Installation extended down times for the instrument (16 working days after the repair/check-up of the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel. In the event of extended down times for the instrument is back in operation.
errors and fault analysisEnvironmental operating conditionsPhilippine environmental conditions: at least to to 40 °C Temperature, 40% to 90% RHB. ACCESORIES, SPARES AND CONSUMABLESAny other accessories require to be quoted to make the system fully operational must be included in the offer.C. POWER SUPPLYC1 Power input should be 220V - 230V, 60HzC.2 Automatic Voltage Regulator (AVR) compatible with the unitWARANTY AND AFTER SALES SERVICESD.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMechDuring warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.D.2 Provision of a qualification package. Package must include Installation qualification (QQ) and Performance qualification (QQ) whichever is applicableD.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. the amount of time elapsed until the structure is back in operation.D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be houred of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be norde ad continued by the new distri
conditions 15 °C to 40 °C Temperature, 40% to 90% RH B. ACCESORIES, SPARES AND CONSUMABLES Any other accessories require to be quoted to make the system fully operational must be included in the offer. C. POWER SUPPLY C.1 Power input should be 220V - 230V, 60Hz C.2 Automatic Voltage Regulator (AVR) compatible with the unit WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel.), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of
B. ACCESORIES, SPARES AND CONSUMABLES Any other accessories require to be quoted to make the system fully operational must be included in the offer. C. POWER SUPPLY C1 Power input should be 220V - 230V, 60Hz C2 Automatic Voltage Regulator (AVR) compatible with the unit WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (IQ), Operational after 15 working days from the time the instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service, personnel. In the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distr
Any other accessories require to be quoted to make the system fully operational must be included in the offer. C. POWER SUPPLY C1 Power input should be 220V - 230V, 60Hz C2 Automatic Voltage Regulator (AVR) compatible with the unit WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time time te instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a tr
operational must be included in the offer.C. POWER SUPPLYC.1 Power input should be 220V - 230V, 60HzC.2 Automatic Voltage Regulator (AVR) compatible with the unitWARRANTY AND AFTER SALES SERVICESD.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMechDuring warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicableD.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranty:D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranty:D.5 In case there is a transfer of sole-distributor.E. TESTING AND ACCEPTANCE
C. POWER SUPPLY C.1 Power input should be 220V - 230V, 60Hz C.2 Automatic Voltage Regulator (AVR) compatible with the unit WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shalb honored and continued by the new distributor. E. TESTING AND ACCEPTANCE E
C2 Automatic Voltage Regulator (AVR) compatible with the unit WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shalb be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE E. TESTING AND ACCEPTANCE
WARRANTY AND AFTER SALES SERVICES D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (IQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor shall ensure that the warrantise and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
D.1 A minimum of one (1) year warranty on parts and services of Refrigerated Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMechDuring warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicableD.3 Instrument should be attended for repair by trained service personnel within fitteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation.D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be honored and continued by the new distributor.E. TESTING AND ACCEPTANCE
Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMechDuring warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program.D.2 Provision of a qualification package. Package must include Installation Qualification (PQ) whichever is applicableD.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation.D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor.E. TESTING AND ACCEPTANCE
Incubator system. Commencing after successful installation by Supplier and its acceptance by PHilMech During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
During warranty period, Supplier must repair or replace any system component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
component on site at no charge to PHilMech. Labor and travel must be included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
included in the warranty program. D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
D.2 Provision of a qualification package. Package must include Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ) whichever is applicableD.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation.D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor.E. TESTING AND ACCEPTANCE
Qualification (PQ) whichever is applicable D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
 D.3 Instrument should be attended for repair by trained service personnel within fifteen (15) working days from the time fault is reported by the user. The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
within fifteen (15) working days from the time fault is reported by the user.The equipment shall be made fully operational after 15 working days from thetime the instrument is attended by the service personnel. In the event ofextended down times for the instrument (16 working days after therepair/check-up of the service personnel), the warranty will be extended bythe amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questionsconcerning maintenance, service, troubleshooting and repairs during theentire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributorshall ensure that the warranties and after sales services given to PHilMechshall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
The equipment shall be made fully operational after 15 working days from the time the instrument is attended by the service personnel. In the event of extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
extended down times for the instrument (16 working days after the repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
repair/check-up of the service personnel), the warranty will be extended by the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
the amount of time elapsed until the instrument is back in operation. D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
D.4 The Supplier must provide technical support, free of charge for questions concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty.D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor.E. TESTING AND ACCEPTANCE
concerning maintenance, service, troubleshooting and repairs during the entire period of the warranty. D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
D.5 In case there is a transfer of sole-distributor rights, the former distributor shall ensure that the warranties and after sales services given to PHilMech shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
shall ensure that the warranties and after sales services given to PHilMechshall be honored and continued by the new distributor.E. TESTING AND ACCEPTANCE
shall be honored and continued by the new distributor. E. TESTING AND ACCEPTANCE
E. TESTING AND ACCEPTANCE
E.1 Acceptance testing will occur after completion of the demonstration of
operation and after the supplier states the system is installed and ready to
perform in compliance with specifications
E.2 Includes calibration by ISO 17025 Calibration Service Provider prior to acceptance with 3 set points set by the end-user (with Calibration Certificate)
F. DOCUMENTATION
F.1 Training certificate of Service Personnel related to the installation,
equipment maintenance and troubleshooting must be provided.
F.2 Manufacturer's certificate of compliance and inspection must be provided.
F.3 One (1) original copy of operating and service manuals in English.
F.4 IQ, OQ, and PQ protocols whichever is applicable. The protocols must
provide the necessary information and test procedures required to prove that
the instrument meets the specified requirements. F.5 Submission of IQ, OQ, and PQ (whichever is applicable) reports after
completion of qualification. Test data (raw data) and calibration certificates of
the calibrating instruments must be attached. Certificate of qualification must
also be issued, if applicable.
G. TRAINING
G.1 The training should cover lectures on the theory and principles of the

``Read and accepted as part of the contract"

instrument and its applications. Troubleshooting, maintenance, and service, including the verification of equipment's performance should also be discussed. Lecture and hands-on training should be conducted at the PHilMech Office for at least five (5) participants. The training shall be provided at no cost to the PHilMech	
H. OTHER REQUIREMENTS	
H.1 The brand has been in the market for at least two years.	
H.2 Supplier must be in the business of supplying Incubator and related equipment for at least two years	
H.3 The supplier must have at least one (1) local installation of Refrigerated	
(or Cooled) Incubator that can be viewed and demo to the end user, if	
necessary.	
Note: Upon submission of bid, the bidder must submit a list of companies with	
installed units. Contact person and contact details (telephone number/mobile	
number and e-mail address) for each listed company must also be submitted.	
I. DELIVERY PERIOD	
I.1 Not more than one hundred twenty (120) calendar days upon receipt of	
Notice to Proceed. To be delivered at Food Protection Division, PHilMech,	
CLSU Compound, Science City of Muñoz, Nueva Ecija.	

I hereby certify that the Statement of Compliance to the Foregoing Technical Specifications are true and correct, otherwise, if found false either during the bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Signature over Printed Name of Bidder

Position / Designation

(Name, Address & Tel. No. of Company)

``Read and accepted as part of the contract"

		Statement of Compliance
	<section-header><section-header></section-header></section-header>	Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.
	ation, Commissioning and Testing of Rotary Shaker	
complete with all necess	ary parts and accessories	
A.1 Technical Specificat		
Parameters	Specifications	
Design	shaker orbital motion	
Speed Range	at least 10 - 250 rpm or better	
Maximum Load	at least 7 kg or higher	
Timer	at least 50 hr 59 min or higher	
Display	Digital (LED or LCD)	
Flask Capacity	at least 250 mL to 500 mL or higher	
Environmental	Philippine environmental condition or better:	
operating conditions	25°C to 31°C Temperature, 70% to 90% RH	
B. ACCESORIES, SPARE		
-	equire to be quoted to make the system fully	
operational must be inc	iuded in the offer.	
C. POWER SUPPLY	h- 2201/ 2201/ (011-	
C.1 Power input should		
	Regulator (AVR) compatible with the unit	
D. WARRANTY, SERVI		
D.1 A minimum of one (1) year warranty on parts and services of Rotary Shaker. Commencing after successful installation by Supplier and its acceptance by PHilMech		
	riod, Supplier must repair or replace any system o charge to PHilMech. Labor and travel must be y program.	

``Read and accepted as part of the contract"

D.2 Provision of a qualification package. Package must include Installation	
Qualification (IQ), Operational Qualification (OQ) and Performance	
Qualification (PQ) whichever is applicable	
D.3 The Supplier must provide technical support, free of charge for	
questions concerning maintenance, service, troubleshooting and repairs	
during the entire period of the warranty.	
D.4 Instrument should be attended for repair by trained service personnel	
within 15 working days from the time fault is reported by the user.	
The equipment shall be made fully operational after 15 working days from	
the time the instrument is attended by the service personnel. In the event	
of extended down times for the instrument (16 working days after the	
repair/check-up of the service personnel), the warranty will be extended	
by the amount of time elapsed until the instrument is back in operation.	
D.6 In case there is a transfer of sole-distributor rights, the former	
distributor shall ensure that the warranties and after sales services given	
to PHilMech shall be honored and continued by the new distributor.	
E. TESTING AND ACCEPTANCE	
E.1 Acceptance testing will occur after completion of the demonstration of	
operation and after the supplier states the system is installed and ready to	
perform in compliance with specifications	
F. DOCUMENTATION	
F.1 Training certificate of Service Personnel related to the installation,	
equipment maintenance and troubleshooting must be provided.	
F.2 Manufacturer's certificate of compliance and inspection must be	
provided.	
F.3 One (1) original copy of operating and service manuals in English.	
F.4 IQ, OQ, and PQ protocols whichever is applicable. The protocols must	
provide the necessary information and test procedures required to prove	
that the instrument meets the specified requirements.	
F.5 Submission of IQ, OQ, and PQ (whichever is applicable) reports after	
completion of qualification. Certificate of qualification must also be issued,	
if applicable	
G. TRAINING	
G.1 The training should cover lectures on the theory and principles of the	
instrument. Troubleshooting, maintenance, and service should also be	
discussed. Lecture and hands-on training should be conducted at the	
PHilMech Office for at least five (5) participants. The training shall be	
provided at no cost to the PHilMech	
H. DELIVERY PERIOD	
I.1 Not more than ninety (90) calendar days upon receipt of Notice to	
Proceed. To be delivered at Food Protection Division, PHilMech, CLSU	
Compound, Science City of Muñoz, Nueva Ecija.	

I hereby certify that the Statement of Compliance to the Foregoing Technical Specifications are true and correct, otherwise, if found false either during the bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Signature over Printed Name of Bidder

Position / Designation

(Name, Address & Tel. No. of Company)

``Read and accepted as part of the contract''

Section VII. Checklist of Technical and Financial Documents

Checklist of Requirements

- □ (a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;
- (b) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
 <u>or</u>
 Original computer of Naturized Bid Security Declaration, and

Original copy of Notarized Bid Securing Declaration; and

 $\hfill\square$ (c) Original of duly signed and accomplished Quotation Form

Quotation Form (Lot No.)

Date : _____ Project Identification No. : _____

To: [name and address of Procuring Entity]

We, the undersigned, offer to [supply/deliver/perform] [description of the Goods] for the sum of [total Bid amount in words and figures] or the total calculated bid price. The total bid price includes the cost of all taxes, such as, but not limited to: [specify the applicable taxes, e.g. (i) value added tax (VAT), (ii) income tax, (iii) local taxes, and (iv) other fiscal levies and duties],

If our offer is accepted, we undertake:

- a. to deliver the goods in accordance with the delivery schedule;
- b. to provide a performance security in the form, amounts, and within the times prescribed in RA9184;

Name:
egal capacity:
Signature:
Duly authorized to sign the Bid for and behalf of:
Date:

BID SECURING DECLARATION FORM

REPUBLIC OF THE PHILIPPINES) CITY OF) S.S.

x-----x

BID SECURING DECLARATION SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT

[PHilMech-Goods-23-11-29-NP)]

Lot no/s. being bid:	Name and Signature of Authorized Representative:

To: [Insert name and address of the Procuring Entity]

I/We², the undersigned, declare that:

- I/We understand that, according to your conditions, bids must be supported by 1. a Bid Security, which may be in the form of a Bid-Securing Declaration.
- 2. I/We accept that: (a) I/we will be automatically disgualified from bidding for any contract with any procuring entity for a period of two (2) years upon receipt of your Blacklisting order; and, (b) I/we will pay the applicable fine provided under Section 6 of the Guidelines on the Use of Bid Securing Declaration, within fifteen (15) days from receipt of the written demand by the procuring entity for the commission of acts resulting to the enforcement of the bid securing declaration under Sections 23.1(b), 34.2, 40.1 and 69.1, except 69.1(f), of the IRR of RA 9184; without prejudice to other legal action the government may undertake.
- 3. I/We understand that this Bid Securing Declaration shall cease to be valid on the following circumstances:
 - Upon expiration of the bid validity period, or any extension thereof (a) pursuant to your request;
 - I am/we are declared ineligible or post-disqualified upon receipt of (b) your notice to such effect, and (i) I/we failed to timely file a request for reconsideration or (ii) I/we filed a waiver to avail of said right;

² Select one and delete the other. Adopt the same instruction for similar terms throughout the document.

(c) I am/we are declared the bidder with the Lowest Calculated Responsive Bid, and I/we have furnished the performance security and signed the Contract.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this _____ day of [month] [year] at [place of execution].

[Insert NAME OF BIDDER'S AUTHORIZED REPRESENTATIVE] [Insert Signatory's Legal Capacity] Affiant

SUBSCRIBED AND SWORN to before me this _____ day of [month] [year] at [place of execution], Philippines. Affiant/s is/are personally known to me and was/were identified by me through competent evidence of identity as defined in the 2004 Rules on Notarial Practice (A.M. No. 02-8-13-SC). Affiant/s exhibited to me his/her [insert type of government identification card used], with his/her photograph and signature appearing thereon, with no. ______ and his/her Community Tax Certificate No. ______ issued on _____ at ____.

Witness my hand and seal this ____ day of [month] [year].

NAME OF NOTARY PUBLIC

Serial No. of Commission	
Notary Public for	_until
Roll of Attorneys No.	
PTR No [date iss	ued], [place issued]
IBP No [date iss	ued], [place issued]

Doc. No	
Page No.	
Book No.	
Series of	

